

Horse SOD1 (Superoxide dismutase) ELISA Kit | SOD1 elisa kit

Horse SOD1 (Superoxide dismutase) ELISA Kit

CATALOG #
AAA248281

REACTIVITY
Horse

PRODUCT SYNONYMS

SOD1 (Superoxide dismutase); Horse SOD1 (Superoxide dismutase) ELISA Kit; SOD1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Horse
Specificity	This assay has high sensitivity and excellent specificity for detection of SOD1. No significant cross-reactivity or interference between SOD1 and analogues was observed.
Sequence Length	204
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.781-50ng/ml
Sensitivity	0.469ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level SOD1 were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level SOD1 were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard operation of a SOD1 ELISA Kit are listed below. This standard curve was generated at our lab for demonstration purpose only. Users shall obtain standard curve as per experiment by themselves. (N/A=not applicable)

X	ng/ml	0	0.781	1.562	3.125	6.25	12.5	25	50
Y	OD450	0.11	0.4	0.652	1.038	1.451	1.774	2.039	2.239

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for SOD1 elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-SOD1 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-SOD1 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the SOD1 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of SOD1 can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	217184
Molecular Weight	22,905 Da
NCBI Official Full Name	superoxide dismutase
UniProt Protein Name	Superoxide dismutase [Mn]
UniProt Gene Name	sodA

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/horse-elisa-kits/sod1-superoxide-dismutase/248281>