

Human C6 (Complement C6) ELISA Kit | C6 elisa kit

Human C6 (Complement C6) ELISA Kit

CATALOG #
AAA248355

REACTIVITY
Human


PRODUCT SYNONYMS

C6 (Complement C6); Human C6 (Complement C6) ELISA Kit; C6 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Sequence Length	934
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.625-40ng/ml
Sensitivity	0.375ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for C6 elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti- C6 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-C6 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the C6 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of C6 can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #

146345396

NCBI GeneID	729
UniProt Accession #	P13671
Molecular Weight	104,786 Da
NCBI Official Full Name	Complement component C6
NCBI Official Synonym Full Names	complement C6
NCBI Official Symbol	C6
NCBI Protein Information	complement component C6
UniProt Protein Name	Complement component C6
UniProt Gene Name	C6

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/c6-complement-c6/248355>