

Human YBX1 (Nuclease-sensitive element-binding protein 1) ELISA Kit | YBX1 elisa kit

Human YBX1 (Nuclease-sensitive element-binding protein 1) ELISA Kit

CATALOG #
AAA248793

REACTIVITY
Human

PRODUCT SYNONYMS

YBX1 (Nuclease-sensitive element-binding protein 1); Human YBX1 (Nuclease-sensitive element-binding protein 1) ELISA Kit; YBX1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of YBX1. No significant cross-reactivity or interference between YBX1 and analogues was observed.
Samples	serum, plasma, tissue homogenates and other biological fluids
Assay Type	Quantitative Sandwich
Detection Range	0.313-20 ng/ml
Sensitivity	<0.188 ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level YBX1 were tested 20 times on one plate, respectively. Intra-Assay: CV < 8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level YBX1 were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/mean X 100. Inter-Assay: CV < 10%
Preparation and Storage	Store entire kit at 2-8°C for short-term. For longer-term, please store the microplate & standard at -20°C, while the remaining reagents can be stored at 2-8°C

IMAGES

Typical Data & Standard Curve

Results of a typical standard operation of a YBX1 ELISA Kit are listed below. This standard curve was generated at our lab for demonstration purpose only. Users shall obtain standard curve as per experiment by themselves. (N/A not applicable)

X	ng/ml	0	0.312	0.625	1.25	2.5	5	10	20
Y	OD450	0.014	0.224	0.392	0.662	1.112	1.495	1.748	1.971

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for YBX1 elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-YBX1 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-YBX1 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the YBX1 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of YBX1 can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.