

Human ATRN (Attractin) ELISA Kit | ATRN elisa kit

Human ATRN (Attractin) ELISA Kit

CATALOG #
AAA248797

REACTIVITY
Human

PRODUCT SYNONYMS

ATR N (Attractin); Human ATR N (Attractin) ELISA Kit; ATR N elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of ATRN. No significant cross-reactivity or interference between ATRN and analogues was observed.
Sequence Length	1272
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	15.625-1000pg/ml
Sensitivity	9.375pg/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level ATRN were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level ATRN were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard operation of a ATRN ELISA Kit are listed below. This standard curve was generated at our lab for demonstration purpose only. Users shall obtain standard curve as per experiment by themselves. (N/A: not applicable)

X	pg/ml	0	15.625	31.25	62.5	125	250	500	1000
Y	OD450	0.096	0.153	0.22	0.28	0.36	0.738	1.221	2.367

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ATRN elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-ATRN antibody was pre-coated onto 96-well plates. And the biotin conjugated anti- ATRN antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the ATRN amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of ATRN can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	75516722
NCBI GeneID	8455
Molecular Weight	133,702 Da
NCBI Official Full Name	Attractin
NCBI Official Synonym Full Names	attractin
NCBI Official Symbol	ATRN
NCBI Official Synonym Symbols	MGCA; DPPT-L
NCBI Protein Information	attractin
UniProt Protein Name	Attractin
UniProt Gene Name	ATRN
UniProt Synonym Gene Names	KIAA0548; MGCA

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/atrn-attractin/248797>