

Human A3/ADORA3 (Adenosine A3 Receptor) ELISA Kit | A3/ADORA3 elisa kit

Human A3/ADORA3 (Adenosine A3 Receptor) ELISA Kit

CATALOG #
AAA248807

REACTIVITY
Human

PRODUCT SYNONYMS

A3/ADORA3 (Adenosine A3 Receptor); Human A3/ADORA3 (Adenosine A3 Receptor) ELISA Kit; A3/ADORA3 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of A3/ADORA3. No significant cross-reactivity or interference between A3/ADORA3 and analogues was observed.
Sequence Length	318
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.156-10ng/ml
Sensitivity	0.094ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level A3/ADORA3 were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level A3/ADORA3 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard run of a A3/ADORA3 ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A/not applicable)

X	ng/ml	0	0.156	0.312	0.625	1.25	2.5	5	10
Y	OD450	0.001	0.076	0.128	0.275	0.526	0.888	1.992	1.984

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for A3/ADORA3 elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-A3/ADORA3 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-A3/ADORA3 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the A3/ADORA3 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of A3/ADORA3 can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	718550823
NCBI GeneID	704295
NCBI Accession #	NP_001289711.1
NCBI GenBank Nucleotide #	NM_001302782.1
NCBI Official Full Name	adenosine A3 receptor
NCBI Official Symbol	ADORA3
NCBI Protein Information	adenosine A3 receptor

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/a3-adora3-adenosine-a3-receptor/248807>