

Rat IL-12A (Interleukin 12 p35) ELISA Kit | IL-12A elisa kit

Rat IL-12A (Interleukin 12 p35) ELISA Kit

CATALOG #
AAA249008

REACTIVITY
Rat

PRODUCT SYNONYMS

IL-12A (Interleukin 12 p35); Rat IL-12A (Interleukin 12 p35) ELISA Kit; IL-12A elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of IL-12A. No significant cross-reactivity or interference between IL-12A and analogues was observed.
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	15.625-1000pg/ml
Sensitivity	9.375pg/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level IL-12A were tested 20 times on one plate, respectively. Intra-Assay: CV<8%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level IL-12A were tested on 3 different plates, 8 replicates in each plate. CV (%) = SD/meanX100. Inter-Assay: CV<10%
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES

Typical Data & Standard Curve

Results of a typical standard operation of a IL-12A ELISA Kit are listed below. This standard curve was generated at our lab for demonstration purpose only. Users shall obtain standard curve as per experiment by themselves. (N/A-not applicable)

X	pg/ml	0	15.625	31.25	62.5	125	250	500	1000
Y	OD450	0.085	0.118	0.121	0.164	0.252	0.47	0.939	1.983

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for IL-12A elisa kit

Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Anti-IL-12A antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-IL-12A antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and washed with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the IL-12A amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of IL-12A can be calculated.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.