

Mouse anti-NMDAR Antibody (anti-N-methyl-D-aspartic Acid receptor Antibody) ELISA Kit | anti-NMDAR elisa kit

Mouse anti-NMDAR Antibody (anti-N-methyl-D-aspartic Acid receptor Antibody) ELISA Kit

CATALOG # AAA250864	REACTIVITY Mouse
-------------------------------	----------------------------


PRODUCT SYNONYMS

anti-NMDAR Antibody (anti-N-methyl-D-aspartic Acid receptor Antibody); Mouse anti-NMDAR Antibody (anti-N-methyl-D-aspartic Acid receptor Antibody) ELISA Kit; anti-NMDAR Antibody; anti-N-methyl-D-aspartic Acid receptor Antibody; anti-NMDAR elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Specificity	Specifically recognize anti-NMDAR Antibody, no obvious cross reaction with other analogues
Assay Type	Sandwich
Samples	Serum, plasma, cell culture supernatant and other biological samples
Detection Range	1.563-100ng/ml
Sensitivity	0.938ng/ml
Intra-assay Precision	Intra-assay Precision: samples with low, medium and high concentration are tested 20 times on same plate.
Inter-assay Precision	Inter-assay Precision: samples with low, medium and high concentration are tested 20 times on three different plates.
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for anti-NMDAR elisa kit	<p>Principle of the Assay: This kit was based on sandwich enzyme-linked immune-sorbent assay technology. Antigen was pre-coated onto the 96-well plate. The biotin conjugated antigen was used as the detection antigen. The standards and pilot samples were added to the wells subsequently. After incubation, unbound conjugates were removed by wash buffer. Then, biotinylated detection antigen was added to bind with anti-NMDAR Antibody conjugated on coated antigen. After washing off unbound conjugates, HRP-Streptavidin was added. After a third washing, TMB substrates were added to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that turned yellow after adding acidic stop solution. Read the O.D. absorbance at 450nm in a microplate reader. The concentration of target antigen in the sample is positively correlated with OD450 and can be calculated by plotting the standard curve.</p>
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

