

Monkey RF-IgG (Rheumatoid Factor IgG) ELISA Kit | RF-IgG elisa kit

Monkey RF-IgG (Rheumatoid Factor IgG) ELISA Kit

CATALOG #
AAA250939

REACTIVITY
Monkey


PRODUCT SYNONYMS

RF-IgG (Rheumatoid Factor IgG); Monkey RF-IgG (Rheumatoid Factor IgG) ELISA Kit; RF-IgG; Rheumatoid Factor IgG; RF-IgG elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Monkey
Specificity	Specifically recognize RF-IgG, no obvious cross reaction with other analogues
Samples	Serum, plasma, cell culture supernatant and other biological samples
Detection Range	3.125-200IU/ml
Sensitivity	1.875IU/ml
Intra-assay Precision	Intra-assay Precision: samples with low, medium and high concentration are tested 20 times on same plate.
Inter-assay Precision	Inter-assay Precision: samples with low, medium and high concentration are tested 20 times on three different plates.
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for RF-IgG elisa kit

Principle of the Assay: This kit was based on indirect enzyme-linked immune-sorbent assay technology. Antigen was pre-coated onto the 96-well plate. The biotin conjugated antibody was used as the detection antibody. The standards and pilot samples were added to the wells subsequently. After incubation, unbound conjugates were removed by wash buffer. Then, biotinylated detection antibody was added to wells. After washing off unbound conjugates, HRP-Streptavidin was added. After a third washing, TMB substrates were added to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product th at turned yellow after adding acidic stop solution. Read the O.D. absorbance at 450nm in a microplate reader. The concentration of target antigen in the sample is positively correlated with OD450 and can be calculated by plotting the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

