

Human FcεRII (Receptor II For The Fc Region Of Immunoglobulin E) ELISA Kit | FcεRII elisa kit

Human FcεRII (Receptor II For The Fc Region Of Immunoglobulin E) ELISA Kit

CATALOG # AAA269883	REACTIVITY Human
-------------------------------	----------------------------


PRODUCT SYNONYMS

FcεRII (Receptor II For The Fc Region Of Immunoglobulin E); Human FcεRII (Receptor II For The Fc Region Of Immunoglobulin E) ELISA Kit; CD23; FCER2; BLAST-2; CLEC4J; FCE2; IGEBF; C-type lectin domain family 4 member J; Immunoglobulin E-binding factor; Lymphocyte IgE receptor; FcεRII elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	serum, plasma, tissue homogenates and other biological fluids
Assay Type	Sandwich
Detection Range	0.32-20 ng/mL
Sensitivity	0.115 ng/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for FcεRII elisa kit	Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Receptor II For The Fc Region Of Immunoglobulin E(FcεRII). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Receptor II For The Fc Region Of Immunoglobulin E(FcεRII). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Receptor II For The Fc Region Of Immunoglobulin E(FcεRII), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Receptor II For The Fc Region Of Immunoglobulin E(FcεRII) in the samples is then determined by comparing the OD of the samples to the standard curve.
Product Categories/Family for FcεRII elisa kit	CD & Adhesion molecule; Infection immunity; Rheumatology

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

