

Dog TNFa (Tumor Necrosis Factor Alpha) ELISA Kit | TNFa elisa kit

Dog TNFa (Tumor Necrosis Factor Alpha) ELISA Kit

CATALOG #
AAA269890

REACTIVITY
Dog


PRODUCT SYNONYMS

TNFa (Tumor Necrosis Factor Alpha); Dog TNFa (Tumor Necrosis Factor Alpha) ELISA Kit; TNF Alpha; DIF; TNF-A; TNFSF2; Cachectin; Tumor Necrosis Factor Ligand Superfamily Member 2; TNFa elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Dog
Samples	serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Assay Type	Sandwich
Detection Range	15.63-1000 pg/mL
Sensitivity	6.1 pg/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TNFa elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Tumor Necrosis Factor Alpha(TNFa). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Tumor Necrosis Factor Alpha(TNFa). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Tumor Necrosis Factor Alpha(TNFa), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Tumor Necrosis Factor Alpha(TNFa) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for TNFa elisa kit

Cytokine; Tumor immunity; Infection immunity

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.