

Rat AQP2 (Aquaporin 2, Collecting Duct) ELISA Kit | AQP2 elisa kit

Rat AQP2 (Aquaporin 2, Collecting Duct) ELISA Kit

CATALOG #
AAA270126

REACTIVITY
Rat


PRODUCT SYNONYMS

AQP2 (Aquaporin 2; Collecting Duct); Rat AQP2 (Aquaporin 2; Collecting Duct) ELISA Kit; AQP-CD; WCH-CD; ADH water channel; Aquaporin-CD; Collecting duct water channel protein; Water channel protein for renal collecting duct; AQP2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Samples	tissue homogenates, cell lysates and other biological fluids
Assay Type	Sandwich
Detection Range	0.32-20 ng/mL
Sensitivity	0.122 ng/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AQP2 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Aquaporin 2, Collecting Duct(AQP2). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Aquaporin 2, Collecting Duct(AQP2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Aquaporin 2, Collecting Duct(AQP2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Aquaporin 2, Collecting Duct(AQP2) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for AQP2 elisa kit

Metabolic pathway; Kidney biomarker

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/aqp2-aquaporin-2-collecting-duct/270126>