

Human XRCC6 (X-Ray Repair Cross Complementing 6) ELISA Kit | XRCC6 elisa kit

Human XRCC6 (X-Ray Repair Cross Complementing 6) ELISA Kit

CATALOG #
AAA270520

REACTIVITY
Human


PRODUCT SYNONYMS

XRCC6 (X-Ray Repair Cross Complementing 6); Human XRCC6 (X-Ray Repair Cross Complementing 6) ELISA Kit; Ku70; CTC75; CTCBF; G22P1; ML8; TLAA; Lupus Ku Autoantigen Protein P70; ATP-Dependent DNA Helicase 2 Subunit 1; CTC box-binding factor 75 kDa; Thyroid-lupus autoantigen; XRCC6 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	tissue homogenates, cell lysates and other biological fluids
Assay Type	Sandwich
Detection Range	0.32-20 ng/mL
Sensitivity	0.124 ng/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for XRCC6 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to X-Ray Repair Cross Complementing 6(XRCC6). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to X-Ray Repair Cross Complementing 6(XRCC6). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain X-Ray Repair Cross Complementing 6(XRCC6), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of X-Ray Repair Cross Complementing 6(XRCC6) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for XRCC6 elisa kit

[Signal transduction](#); [Developmental science](#)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/xrcc6-x-ray-repair-cross-complementing-6/270520>