

Human CYFRA21-1 (Cytokeratin Fragment Antigen 21-1) ELISA Kit | CYFRA21-1 elisa kit

Human CYFRA21-1 (Cytokeratin Fragment Antigen 21-1) ELISA Kit

CATALOG #
AAA270550

REACTIVITY
Human


PRODUCT SYNONYMS

CYFRA21-1 (Cytokeratin Fragment Antigen 21-1); Human CYFRA21-1 (Cytokeratin Fragment Antigen 21-1) ELISA Kit; CYFRA21-1; CYFRA21-1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Human CYFRA21-1. No significant cross-reactivity or interference between Human CYFRA21-1 and analogues was observed.
Assay Type	Sandwich
Detection Range	0.32-20 ng/mL
Sensitivity	0.117 ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (Precision between assays)? CV%<10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CYFRA21-1 elisa kit	Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Cytokeratin Fragment Antigen 21-1(CYFRA21-1). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Cytokeratin Fragment Antigen 21-1(CYFRA21-1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Cytokeratin Fragment Antigen 21-1(CYFRA21-1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm $\pm$ 10nm. The concentration of Cytokeratin Fragment Antigen 21-1(CYFRA21-1) in the samples is then determined by comparing the OD of the samples to the standard curve.
Product Categories/Family for CYFRA21-1 elisa kit	Tumor immunity; Dermatology

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

