

Rat EDN1 (Endothelin 1) ELISA Kit | EDN1 elisa kit

Rat EDN1 (Endothelin 1) ELISA Kit

CATALOG #
AAA270915

REACTIVITY
Rat


PRODUCT SYNONYMS

EDN1 (Endothelin 1); Rat EDN1 (Endothelin 1) ELISA Kit; ET1; PPET1; Preproendothelin-1; Big endothelin-1; EDN1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of Rat EDN1. No significant cross-reactivity or interference between Rat EDN1 and analogues was observed.
Assay Type	Sandwich
Detection Range	7.82-500 pg/mL
Sensitivity	2.71 pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (Precision between assays)? CV%<10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for EDN1 elisa kit

Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Endothelin 1(EDN1). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Endothelin 1(EDN1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Endothelin 1(EDN1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm± 10nm. The concentration of Endothelin 1(EDN1) in the samples is then determined by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/edn1-endothelin-1/270915>