

Rabbit TIMP4 (Tissue Inhibitors Of Metalloproteinase 4) ELISA Kit | TIMP4 elisa kit

Rabbit TIMP4 (Tissue Inhibitors Of Metalloproteinase 4) ELISA Kit

CATALOG #
AAA271229

REACTIVITY
Rabbit


PRODUCT SYNONYMS

TIMP4 (Tissue Inhibitors Of Metalloproteinase 4); Rabbit TIMP4 (Tissue Inhibitors Of Metalloproteinase 4) ELISA Kit; TIMP Metalloproteinase Inhibitor 4; Tissue Inhibitor of Metalloproteinases 4; TIMP4 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Samples	Serum, plasma and other biological fluids.
Assay Type	Sandwich
Detection Range	78.13-5000 pg/mL
Sensitivity	27 pg/mL
Intra-assay Precision	CV%<8%
Inter-assay Precision	CV%<10%
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TIMP4 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Tissue Inhibitors Of Metalloproteinase 4(TIMP4). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Tissue Inhibitors Of Metalloproteinase 4(TIMP4). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Tissue Inhibitors Of Metalloproteinase 4(TIMP4), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Tissue Inhibitors Of Metalloproteinase 4(TIMP4) in the samples is then determined by comparing the OD of the samples to the standard curve.

Product Categories/Family for TIMP4 elisa kit

[Tumor immunity](#); [Cardiovascular biology](#)

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.