

## Mouse CX43 (Connexin 43) ELISA Kit | CX43 elisa kit

Mouse CX43 (Connexin 43) ELISA Kit

CATALOG #  
**AAA271267**

REACTIVITY  
**Mouse**


### PRODUCT SYNONYMS

CX43 (Connexin 43); Mouse CX43 (Connexin 43) ELISA Kit; ODDD; GJAL; GJA1; GJ-A1; Gap junction 43 kDa heart protein; Gap Junction Protein Alpha 1; Oculodentodigital Dysplasia; Syndactyly Type III; CX43 elisa kit

### PRODUCT SPECIFICATIONS

|                         |                                                                  |
|-------------------------|------------------------------------------------------------------|
| Reactivity              | Mouse                                                            |
| Samples                 | Tissue homogenates, cell lysates and other biological fluids     |
| Assay Type              | Sandwich                                                         |
| Detection Range         | 0.16-10 ng/mL                                                    |
| Sensitivity             | 0.065 ng/mL                                                      |
| Intra-assay Precision   | CV%<8%                                                           |
| Inter-assay Precision   | CV%<10%                                                          |
| Preparation and Storage | Store at 2-8 degree C. Stable for 6 months from date of receipt. |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for CX43 elisa kit

Principle of the assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Connexin 43(CX43). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Connexin 43(CX43). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Connexin 43(CX43), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of Connexin 43(CX43) in the samples is then determined by comparing the OD of the samples to the standard curve.

#### Product Categories/Family for CX43 elisa kit

Cardiovascular biology; Neuro science

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)  
Online datasheet: <https://stage.aaabiotech.com/mouse-elisa-kits/cx43-connexin-43/271267>