

## Human ADAM15 (A Disintegrin And Metalloprotease 15) ELISA Kit | ADAM15 elisa kit

Human ADAM15 (A Disintegrin And Metalloprotease 15) ELISA Kit

CATALOG #  
**AAA271724**

REACTIVITY  
**Human**


### PRODUCT SYNONYMS

ADAM15 (A Disintegrin And Metalloprotease 15); Human ADAM15 (A Disintegrin And Metalloprotease 15) ELISA Kit; MDC15; Metargidin; Metalloproteinase-like; disintegrin-like; and cysteine-rich protein 15; Metalloprotease RGD disintegrin protein; ADAM15 elisa kit

### PRODUCT SPECIFICATIONS

|                         |                                                                                                                                                                                           |
|-------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Reactivity              | Human                                                                                                                                                                                     |
| Specificity             | This assay has high sensitivity and excellent specificity for detection of Human ADAM15. No significant cross-reactivity or interference between Human ADAM15 and analogues was observed. |
| Assay Type              | Sandwich                                                                                                                                                                                  |
| Detection Range         | 0.16-10 ng/mL                                                                                                                                                                             |
| Sensitivity             | 0.055 ng/mL                                                                                                                                                                               |
| Intra-assay Precision   | Intra-assay Precision (Precision within an assay)? CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.                    |
| Inter-assay Precision   | Inter-assay Precision (Precision between assays)? CV%<10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.                     |
| Preparation and Storage | Store at 2-8 degree C. Stable for 6 months from date of receipt.                                                                                                                          |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for ADAM15 elisa kit

Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to A Disintegrin And Metalloprotease 15(ADAM15). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to A Disintegrin And Metalloprotease 15(ADAM15). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain A Disintegrin And Metalloprotease 15(ADAM15), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of A Disintegrin And Metalloprotease 15(ADAM15) in the samples is then determined by comparing the OD of the samples to the standard curve.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | [contact@aaabiotech.com](mailto:contact@aaabiotech.com)

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/adam15-a-disintegrin-and-metalloprotease-15/271724>