

Human Bcl11A (B-Cell CLL/Lymphoma 11A) ELISA Kit | Bcl11A elisa kit

Human Bcl11A (B-Cell CLL/Lymphoma 11A) ELISA Kit

CATALOG #
AAA272793

REACTIVITY
Human


PRODUCT SYNONYMS

Bcl11A (B-Cell CLL/Lymphoma 11A); Human Bcl11A (B-Cell CLL/Lymphoma 11A) ELISA Kit; BCL11A-L; BCL11A-S; BCL11A-XL; CTIP1; EVI9; Ecotropic Viral Integration Site 9; COUP-TF-interacting protein 1; Ecotropic viral integration site 9 homolog; Zinc finger protein 856; Bcl11A elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Human Bcl11A. No significant cross-reactivity or interference between Human Bcl11A and analogues was observed.
Assay Type	Sandwich
Detection Range	0.32-20 ng/mL
Sensitivity	0.114 ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (Precision between assays)? CV%<10%. Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Preparation and Storage	Store at 2-8 degree C. Stable for 6 months from date of receipt.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Bcl11A elisa kit

Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to B-Cell CLL/Lymphoma 11A(Bcl11A). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to B-Cell CLL/Lymphoma 11A(Bcl11A). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain B-Cell CLL/Lymphoma 11A(Bcl11A), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of B-Cell CLL/Lymphoma 11A(Bcl11A) in the samples is then determined by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/bcl11a-b-cell-cll-lymphoma-11a/272793>