

Chicken COL7 (Collagen Type VII) ELISA Kit | COL7 elisa kit

Chicken COL7 (Collagen Type VII) ELISA Kit

CATALOG #
AAA277127

REACTIVITY
Chicken


PRODUCT SYNONYMS

COL7 (Collagen Type VII); Chicken COL7 (Collagen Type VII) ELISA Kit; COL7-A1; COL7A1; EBD1; EBDCT; EBR1; Long-chain collagen; Epidermolysis Bullosa; Dystrophic; Dominant And Recessive; Collagen Alpha-1 (VII)chain; COL7 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Chicken
Specificity	This assay has high sensitivity and excellent specificity for detection of Chicken COL7. No significant cross-reactivity or interference between Chicken COL7 and analogues was observed.
Assay Type	Sandwich
Detection Range	15.63-1000 pg/mL
Sensitivity	6.1 pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay)? CV%<8% Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.
Inter-assay Precision	Inter-assay Precision (Precision between assays)? CV%<10% Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for COL7 elisa kit

Principle of the Assay: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Collagen Type VII(COL7). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Collagen Type VII(COL7). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Collagen Type VII(COL7), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at awavelength of 450nm ± 10nm. The concentration of Collagen Type VII(COL7) in the samples is then determined by comparing the OD of the samples to the standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/chicken-elisa-kits/col7-collagen-type-vii/277127>