

Human KDM3A (Lysine-specific demethylase 3A) ELISA Kit | KDM3A elisa kit

Human KDM3A (Lysine-specific demethylase 3A) ELISA Kit

CATALOG #
AAA277235


REACTIVITY
Human

PRODUCT SYNONYMS

KDM3A (Lysine-specific demethylase 3A); Human KDM3A (Lysine-specific demethylase 3A) ELISA Kit; JMJD1; JMJD1A; TSGA; JHMD2A; JmjC domain-containing histone demethylation protein 2A; Jumonji domain-containing protein 1A; KDM3A elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Assay Type	Sandwich
Samples	tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	0.16-10ng/mL
Sensitivity	0.054ng/mL
Standard	10ng/mL
Assay Length	3.5h
Precision	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision. Inter-assay Precision (Precision between assays): CV%<10% Three samples of known concentration were tested in forty separate assays to assess inter-assay precision.
Test Principle	The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Lysine-specific demethylase 3A (KDM3A). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Lysine-specific demethylase 3A (KDM3A). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Lysine-specific demethylase 3A (KDM3A), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm +/- 10 nm. The concentration of Lysine-specific demethylase 3A (KDM3A) in the samples is then determined by comparing the OD of the samples to the standard curve.
Standard Curve	Matrices listed below were spiked with certain level of recombinant KDM3A and the recovery rates were calculated by comparing the measured value to the expected amount of KDM3A in samples. Matrix Recovery range Average serum (n=5) 95-107% 101% EDTA plasma (n=5) 90-107% 98% Heparin plasma (n=5) 87-99% 93%
Recovery Rate	The linearity of the kit was assayed by testing samples spiked with appropriate concentration of KDM3A and their serial dilutions. The results were demonstrated by the percentage of calculated concentration to the expected. Matrix 1:21:41:81:16 serum (n=5) 87-98% 92-101% 82-98% 89-102% EDTA plasma (n=5) 91-97% 81-93% 89-103% 82-91% Heparin plasma (n=5) 92-101% 80-94% 86-101% 89-98%



Standard Curve (Sample)

ADDITIONAL INFORMATION

Product Categories/Family for KDM3A
elisa kit

Enzyme & Kinase

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/kdm3a-lysine-specific-demethylase-3a/277235>