

Rat N-methyl-d-aspartic acid, NMDA ELISA Kit | NMDA elisa kit

Rat N-methyl-d-aspartic acid, NMDA ELISA Kit

CATALOG #
AAA43652

REACTIVITY
Rat


PRODUCT SYNONYMS

N-methyl-d-aspartic acid; NMDA; Rat N-methyl-d-aspartic acid; NMDA ELISA Kit; NMDA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Samples	Serum, plasma, cell culture supernates, cell lysates, tissue homogenates
Assay Type	Quantitative Sandwich
Detection Range	0.05ng/ml - 30ng/ml
Sensitivity	0.025ng/ml
Intra-assay Precision	Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision. Intra-Assay: CV<8%
Inter-assay Precision	Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision. CV(%) = SD/mean x 100. Inter-Assay: CV<10%
Preparation and Storage	Store at 2 to 8 degree C for 6 months.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for NMDA elisa kit

Intended Uses: This sandwich kit is for the accurate quantitative detection of Rat N-methyl-d-aspartic acid (also known as NMDA) in serum, plasma, cell culture supernates, cell lysates, tissue homogenates.

Principle of the Assay: This kit is an Enzyme-Linked Immunosorbent Assay (ELISA). The plate has been pre-coated with Rat NMDA antibody. NMDA present in the sample is added and binds to antibodies coated on the wells. And then biotinylated Rat NMDA Antibody is added and binds to NMDA in the sample. Then Streptavidin-HRP is added and binds to the Biotinylated NMDA antibody. After incubation unbound Streptavidin-HRP is washed away during a washing step. Substrate solution is then added and color develops in proportion to the amount of Rat NMDA. The reaction is terminated by addition of acidic stop solution and absorbance is measured at 450 nm.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/n-methyl-d-aspartic-acid-nmda/43652>