

Human Cryptochrome 1 (CRY1) ELISA Kit | CRY1 elisa kit

Human Cryptochrome 1 (CRY1) ELISA Kit

CATALOG #
AAA64238

REACTIVITY
Human


PRODUCT SYNONYMS

Cryptochrome 1 (CRY1); Human Cryptochrome 1 (CRY1) ELISA Kit; Cathepsin E; CRY1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of CRY1. No significant cross-reactivity or interference between CRY1 and analogues was observed.
Sequence Length	586
Assay Type	Quantitative Sandwich
Samples	Human tissue homogenates, cell lysates or other biological fluids
Detection Range	0.312-20ng/mL
Sensitivity	< 0.117ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level CRY1 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level CRY1 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CRY1 elisa kit	Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to CRY1. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to CRY1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain CRY1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of CRY1 in the samples is then determined by comparing the O.D. of the samples to the standard curve. Intended Uses: The kit is a sandwich enzyme immunoassay for the in vitro quantitative measurement of CRY1 in human tissue homogenates, cell lysates or other biological fluids.
--	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	4758072
NCBI GeneID	1407
NCBI Accession #	NP_004066.1
NCBI GenBank Nucleotide #	NM_004075.4
UniProt Accession #	Q16526
Molecular Weight	66,395 Da
NCBI Official Full Name	cryptochrome-1
NCBI Official Synonym Full Names	cryptochrome circadian clock 1
NCBI Official Symbol	CRY1
NCBI Official Synonym Symbols	PHLL1
NCBI Protein Information	cryptochrome-1
UniProt Protein Name	Cryptochrome-1
UniProt Gene Name	CRY1
UniProt Synonym Gene Names	PHLL1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.