

General Asymmetrical Dimethylarginine (ADMA) ELISA Kit | ADMA elisa kit

General Asymmetrical Dimethylarginine (ADMA) ELISA Kit

CATALOG #
AAA65149

REACTIVITY
General


PRODUCT SYNONYMS

Asymmetrical Dimethylarginine (ADMA); General Asymmetrical Dimethylarginine (ADMA) ELISA Kit; N; N-Dimethylarginine; NG; NG-Dimethylarginine dihydrochloride; Asymmetric Dimethylarginine; ADMA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of ADMA. No significant cross-reactivity or interference between ADMA and analogues was observed.
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates or other biological fluids
Assay Type	Quantitative Competitive
Detection Range	15.6-1000ng/mL
Sensitivity	< 4.63ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level ADMA were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level ADMA were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ADMA elisa kit

Intended Uses: The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative measurement of ADMA in general serum, plasma, tissue homogenates, cell lysates, cell culture supernates or other biological fluids.

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to ADMA has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled ADMA and unlabeled ADMA (Standards or samples) with the pre-coated antibody specific to ADMA. After incubation the unbound conjugate is washed off. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of ADMA in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of ADMA in the sample.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/general-elisa-kits/asymmetrical-dimethylarginine-adma/65149>