

General Reverse Triiodothyronine (rT3) ELISA Kit | rT3 elisa kit

General Reverse Triiodothyronine (rT3) ELISA Kit

CATALOG #
AAA65158

REACTIVITY
General


PRODUCT SYNONYMS

Reverse Triiodothyronine (rT3); General Reverse Triiodothyronine (rT3) ELISA Kit; RT3; r-T3; Reverse T3;(2S)-2-Amino-3-[4-(4-Hydroxy-3; 5-Diodophenoxy)-3-Iodophenyl]propanoic Acid; 3; 3'; 5'-Triiodo-L-thyronine; rT3 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of rT3. No significant cross-reactivity or interference between rT3 and analogues was observed.
Assay Type	Quantitative Competitive
Samples	General serum, plasma or other biological fluids
Detection Range	93.7-6000pg/mL
Sensitivity	< 24.32pg/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level rT3 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level rT3 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for rT3 elisa kit

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to rT3 has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled rT3 and unlabeled rT3 (Standards or samples) with the pre-coated antibody specific to rT3. After incubation the unbound conjugate is washed off. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of rT3 in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of rT3 in the sample.

Intended Uses: The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative measurement of rT3 in general serum, plasma or other biological fluids.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/general-elisa-kits/reverse-triiodothyronine-rt3/65158>