

Human Chloride Channel Accessory 1 (CLCA1) ELISA Kit | CLCA1 elisa kit

Human Chloride Channel Accessory 1 (CLCA1) ELISA Kit

CATALOG #
AAA66159

REACTIVITY
Human


PRODUCT SYNONYMS

Chloride Channel Accessory 1 (CLCA1); Human Chloride Channel Accessory 1 (CLCA1) ELISA Kit; CaCC1; CLCRG1; Chloride Channel; Calcium Activated; Family Member 1; Calcium-activated chloride channel protein 1; CLCA1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of CLCA1. No significant cross-reactivity or interference between CLCA1 and analogues was observed.
Sequence Length	914
Samples	Tissue homogenates, cell lysates and other biological fluids.
Assay Type	Sandwich
Detection Range	0.156-10ng/mL
Sensitivity	0.056ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level CLCA1 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level CLCA1 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CLCA1 elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of CLCA1 in human tissue homogenates, cell lysates and other biological fluids. Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to CLCA1. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to CLCA1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain CLCA1, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of CLCA1 in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	110611231
NCBI GeneID	1179
NCBI Accession #	NP_001276.2
NCBI GenBank Nucleotide #	NM_001285.3
UniProt Accession #	A8K7I4; O95151; Q5TDF4; Q9UNF6; Q9UPC6; B2RAV5
Molecular Weight	100,226 Da
NCBI Official Full Name	calcium-activated chloride channel regulator 1
NCBI Official Synonym Full Names	chloride channel accessory 1
NCBI Official Symbol	CLCA1
NCBI Official Synonym Symbols	CACC; GOB5; CACC1; CLCRG1; CaCC-1; hCLCA1; hCaCC-1
NCBI Protein Information	calcium-activated chloride channel regulator 1
UniProt Protein Name	Calcium-activated chloride channel regulator 1
UniProt Gene Name	CLCA1
UniProt Synonym Gene Names	CACC1; hCLCA1; CaCC-1; hCaCC-1

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.