

Human Procollagen Lysine-2-Oxoglutarate-5-Dioxygenase 3 (PLOD3) ELISA Kit | PLOD3 elisa kit

Human Procollagen Lysine-2-Oxoglutarate-5-Dioxygenase 3 (PLOD3) ELISA Kit

CATALOG # AAA67388	REACTIVITY Human
------------------------------	----------------------------


PRODUCT SYNONYMS

Procollagen Lysine-2-Oxoglutarate-5-Dioxygenase 3 (PLOD3); Human Procollagen Lysine-2-Oxoglutarate-5-Dioxygenase 3 (PLOD3) ELISA Kit; LH3; Lysyl hydroxylase 3; PLOD3 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of PLOD3. No significant cross-reactivity or interference between PLOD3 and analogues was observed.
Sequence Length	738
Samples	Tissue homogenates and other biological fluids.
Assay Type	Sandwich
Detection Range	0.312-20ng/mL
Sensitivity	0.106ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level PLOD3 were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level PLOD3 were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PLOD3 elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of PLOD3 in human tissue homogenates and other biological fluids. Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to PLOD3. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to PLOD3. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain PLOD3, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of PLOD3 in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	---

NCBI AND UNIPROT INFORMATION

NCBI GI #	4505891
NCBI GeneID	8985
NCBI Accession #	NP_001075.1
NCBI GenBank Nucleotide #	NM_001084.4
UniProt Accession #	O60568; Q540C3; B2R6W6
Molecular Weight	84,785 Da
NCBI Official Full Name	procollagen-lysine,2-oxoglutarate 5-dioxygenase 3
NCBI Official Synonym Full Names	procollagen-lysine,2-oxoglutarate 5-dioxygenase 3
NCBI Official Symbol	PLOD3
NCBI Official Synonym Symbols	LH3
NCBI Protein Information	procollagen-lysine,2-oxoglutarate 5-dioxygenase 3
UniProt Protein Name	Procollagen-lysine,2-oxoglutarate 5-dioxygenase 3
UniProt Gene Name	PLOD3
UniProt Synonym Gene Names	LH3

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/procollagen-lysine-2-oxoglutarate-5-dioxygenase-3-plod3/67388>