

Human CD82 molecule ELISA Kit | CD82 elisa kit

Human cluster Of differentiation 82, CD82 ELISA Kit

CATALOG #
AAA80077

REACTIVITY
Human


PRODUCT SYNONYMS

CD82 molecule; Human cluster Of differentiation 82; CD82 ELISA Kit; Human cluster Of differentiation 82 (CD82) ELISA Kit; 4F9; C33; GR15; IA4; KAI1; R2; SAR2; ST6; TSPAN27; C33 antigen; CD82 antigen; R2 leukocyte antigen; inducible membrane protein R2; kangai 1 (suppression of tumorigenicity 6; prostate; CD82 antigen (R2 I; CD82 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of Human CD82. No significant cross-reactivity or interference between Human CD82 and analogues was observed.
Sequence Length	242
Detection Range	0.156 ng/ml-10 ng/ml
Sensitivity	0.039 ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV% is less than 8% Three samples of known concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV% is less than 10% Three samples of known concentration were tested in twenty assays to assess.
Detection Wavelength	450 nm
Sample Volume	50-100ul
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CD82 elisa kit

Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to CD82. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated polyclonal antibody preparation specific for CD82 and Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then a TMB (3,3',5,5' tetramethyl-benzidine) substrate solution is added to each well. Only those wells that contain CD82, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. The concentration of CD82 in the samples is then determined by comparing the O.D. of the samples to the standard curve.

NCBI AND UNIPROT INFORMATION

NCBI GI #	67782354
NCBI GeneID	3732
NCBI Accession #	NP_001020015.1
NCBI GenBank Nucleotide #	NM_001024844.1
UniProt Accession #	P27701; Q7Z2D4; Q7Z5N2; D3DQN6; E9PC70
Molecular Weight	29,626 Da
NCBI Official Full Name	CD82 antigen isoform 2
NCBI Official Synonym Full Names	CD82 molecule
NCBI Official Symbol	CD82
NCBI Official Synonym Symbols	R2; 4F9; C33; IA4; ST6; GR15; KAI1; SAR2; TSPAN27
NCBI Protein Information	CD82 antigen; tspan-27; C33 antigen; tetraspanin-27; inducible membrane protein R2; metastasis suppressor Kangai-1; kangai 1 (suppression of tumorigenicity 6, prostate; CD82 antigen (R2 leukocyte antigen, antigen detected by monoclonal and antibody IA4))
UniProt Protein Name	CD82 antigen
UniProt Gene Name	CD82
UniProt Synonym Gene Names	KAI1; SAR2; ST6; TSPAN27; Tspan-27
UniProt Entry Name	CD82_HUMAN

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/cd82-molecule/80077>