

Human Influenza A virus antibody (IgG) ELISA Kit

Human Influenza A virus antibody, IgG ELISA Kit

CATALOG #
AAA80431

REACTIVITY
Human

PRODUCT SYNONYMS

Influenza A virus antibody (IgG); Human Influenza A virus antibody; IgG ELISA Kit; Human Influenza A virus antibody (IgG) ELISA Kit; Influenza A virus antibody (IgG) elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum
Detection Wavelength	450 nm
Sample Volume	50-100ul
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES

TYPICAL DATA		
Test parameter	specification	test result
Positive control	>1.0	1.189
Negative control	<0.2	0.087
Positive rate	10. Positive	100%
Negative rate	10. Negative	100%

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Influenza A virus antibody (IgG) elisa kit

Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with Influenza A virus antigen. Samples are then added to the appropriate microtiter plate wells and incubated. Then add Horseradish Peroxidase (HRP)-conjugated anti-human IgG to each well and incubate. Finally, the substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color change is measured spectrophotometrically at a wavelength of 450 nm +/- 2 nm. Calculate the valence of human influenza A virus IgG antibody in the samples.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/influenza-a-virus-antibody-igg/80431>