

Monkey Testosterone (T) ELISA Kit | T elisa kit

Monkey Testosterone, T ELISA Kit

CATALOG #
AAA80891

REACTIVITY
Monkey


PRODUCT SYNONYMS

Testosterone (T); Monkey Testosterone; T ELISA Kit; Monkey Testosterone (T) ELISA Kit; Monkey Testosterone (T); T elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Monkey
Specificity	This assay has high sensitivity and excellent specificity for detection of monkey testosterone. No significant cross-reactivity or interference between monkey testosterone and analogues was observed.
Samples	Serum, plasma, tissue homogenates, cell culture supernates
Assay Type	Quantitative Competitive
Detection Range	0.1 ng/ml-20 ng/ml
Sensitivity	< 0.05 ng/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV%<15%. Three samples of known concentration were tested twenty times on one plate to assess.
Inter-assay Precision	Inter-assay Precision (Precision between assays): CV%<15%. Three samples of known concentration were tested in twenty assays to assess.
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for T elisa kit

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with goat-anti-rabbit antibody. Standards or samples are added to the appropriate microtiter plate wells with an antibody specific for testosterone and Horseradish Peroxidase (HRP) conjugated testosterone. The competitive inhibition reaction is launched between with HRP labeled testosterone and unlabeled testosterone with the antibody. A substrate solution is added to the wells and the color develops in opposite to the amount of testosterone in the sample. The color development is stopped and the intensity of the color is measured.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/monkey-elisa-kits/testosterone-t/80891>