

Rat beta-crosslaps (bCTx) ELISA Kit | bCTx elisa kit

Rat beta-crosslaps, bCTx ELISA Kit

CATALOG #
AAA80963

REACTIVITY
Rat


PRODUCT SYNONYMS

beta-crosslaps (bCTx); Rat beta-crosslaps; bCTx ELISA Kit; Rat beta-crosslaps (bCTx) ELISA kit; bCTx elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of rat bCTx. No significant cross-reactivity or interference between rat bCTx and analogues was observed.
Samples	Serum, plasma, tissue homogenates
Assay Type	Sandwich
Detection Range	7.8 pg/ml -500 pg/ml.
Sensitivity	The minimum detectable dose of rat bCTx is typically less than 1.95 pg/ml. The sensitivity of this assay, or Lower Limit of Detection (LLD) was defined as the lowest protein concentration that could be differentiated from zero. It was determined the mean O.D value of 20 replicates of the zero standard added by their three standard deviations.
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV%<8%. Three samples of known concentration were tested twenty times on one plate to assess.
Inter-assay Precision	Inter-assay Precision (Precision between assays): CV%<10%. Three samples of known concentration were tested in twenty assays to assess.
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES



ADDITIONAL INFORMATION

Related Product Information for bCTx elisa kit

Principle of the assay: This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for bCTx has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any bCTx present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for bCTx is added to the wells. After washing, avidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of bCTx bound in the initial step. The color development is stopped and the intensity of the color is measured.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/rat-elisa-kits/beta-crosslaps-bctx/80963>