

Rabbit 17-Hydroxycorticosteroids ELISA Kit | 17-OHCS elisa kit

Rabbit 17-Hydroxycorticosteroids ELISA Kit

CATALOG #
AAA81831

REACTIVITY
Rabbit


PRODUCT SYNONYMS

17-Hydroxycorticosteroids; Rabbit 17-Hydroxycorticosteroids ELISA Kit; 17-OHCS elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	2.5-50 nmol/mL
Sensitivity	0.1nmol/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for 17-OHCS elisa kit	For Samples: Cell culture fluid & body fluid & tissue homogenate Serum or blood plasma Intended Uses: This 17-OHCS ELISA kit is intended for laboratory research use only. The stop solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of 17-OHCS in the sample, this 17-OHCS ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus 17-OHCS concentration. The concentration of in the samples is then determined by comparing the O.D. of the samples to the standard curve. Principle of the Assay: The coated well immunoenzymatic assay for the quantitative measurement of 17-OHCS utilizes a multiclonal anti-17-OHCS antibody and an 17-OHCS-HRP conjugate. The assay sample and buffer are incubated together with 17-OHCS-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the 17-OHCS concentration since 17-OHCS from samples and 17-OHCS-HRP conjugate compete for the anti-17-OHCS antibody binding site. Since the number of sites is limited, as more sites are occupied by 17-OHCS from the sample, fewer sites are left to bind 17-OHCS-HRP conjugate. Standards of known 17-OHCS concentrations are run concurrently with the samples being assayed and a standard curve is plotted relating the intensity of the color (Optical Density) to the concentration of 17-OHCS. The 17-OHCS concentration in each sample is interpolated from this standard curve.
---	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/17-ohcs-rabbit-elisa-kits/17-hydroxycorticosteroids/81831>