

Human Gastrointestinalcancer Marker CA199 ELISA Kit | Gastrointestinalcancer Marker CA199 elisa kit

Human Gastrointestinalcancer Marker CA199 ELISA Kit

CATALOG #
AAA82567

REACTIVITY
Human


PRODUCT SYNONYMS

Gastrointestinalcancer Marker CA199; Human Gastrointestinalcancer Marker CA199 ELISA Kit; Gastrointestinalcancer Marker CA199 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Detection Range	1.0-25U/mL
Sensitivity	0.1U/mL
Intended Uses	This CA199 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human CA199. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Gastrointestinalcancer Marker CA199 elisa kit

Principle of the Assay: CA199 ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-CA199 antibody and an CA199-HRP conjugate. The assay sample and buffer are incubated together with CA199-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the CA199 concentration since CA199 from samples and CA199-HRP conjugate compete for the anti-CA199 antibody binding site. Since the number of sites is limited, as more sites are occupied by CA199 from the sample, fewer sites are left to bind CA199-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The CA199 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/human-elisa-kits/gastrointestinalcancer-marker-ca199/82567>