

Human Crosslaps ELISA Kit | Cr elisa kit

Human Crosslaps ELISA Kit

CATALOG #
AAA83785

REACTIVITY
Human


PRODUCT SYNONYMS

Crosslaps; Human Crosslaps ELISA Kit; Crosslaps (Human); Cr elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Sensitivity	0.1 ng/mL.
Intended Uses	This Crosslaps ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human Crosslaps. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Cr elisa kit

Principle of the assay: Crosslaps ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-Crosslaps antibody and an Crosslaps-HRP conjugate. The assay sample and buffer are incubated together with Crosslaps-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the Crosslaps concentration since Crosslaps from samples and Crosslaps-HRP conjugate compete for the anti-Crosslaps antibody binding site. Since the number of sites is limited, as more sites are occupied by Crosslaps from the sample, fewer sites are left to bind Crosslaps-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The Crosslaps concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/cr-human-elisa-kits/crosslaps/83785>