

Mouse Anti-Teichoic acid antibody ELISA Kit | TA elisa kit

Mouse Anti-Teichoic acid antibody ELISA Kit

CATALOG #
AAA84466

REACTIVITY
Mouse


PRODUCT SYNONYMS

Anti-Teichoic acid antibody; Mouse Anti-Teichoic acid antibody ELISA Kit; TA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Mouse
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	25-500ng/mL
Sensitivity	1.0ng/mL
Intended Uses	This ATAb ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Mouse ATAb. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TA elisa kit	For Samples: Cell culture fluid, body fluid, tissue homogenate, serum and blood plasma Intended Uses: This ATAAB ELISA kit is intended for laboratory research use only. The stop solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of ATAAB in the sample, this ATAAB ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus ATAAB concentration. The concentration of in the samples is then determined by comparing the O.D. of the samples to the standard curve. Principle of the Assay: The coated well immunoenzymatic assay for the quantitative measurement of ATAAB utilizes a multiclonal anti-ATAAB antibody and an ATAAB-HRP conjugate. The assay sample and buffer are incubated together with ATAAB-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the ATAAB concentration since ATAAB from samples and ATAAB-HRP conjugate compete for the anti-ATAAB antibody binding site. Since the number of sites is limited, as more sites are occupied by ATAAB from the sample, fewer sites are left to bind ATAAB-HRP conjugate. Standards of known ATAAB concentrations are run concurrently with the samples being assayed and a standard curve is plotted relating the intensity of the color (Optical Density) to the concentration of ATAAB. The ATAAB concentration in each sample is interpolated from this standard curve.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/ta-mouse-elisa-kits/anti-teichoic-acid-antibody/84466>