

Rat Oxidized lowdensity lipoprotein ELISA Kit | OxLDL elisa kit

Rat Oxidized lowdensity lipoprotein ELISA Kit

CATALOG #
AAA86027

REACTIVITY
Rat


PRODUCT SYNONYMS

Oxidized lowdensity lipoprotein; Rat Oxidized lowdensity lipoprotein ELISA Kit; OxLDL elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Samples	Serum, Plasma, Cell Culture Supernatants, Body Fluid And Tissue Homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 U/mL.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for OxLDL elisa kit

Intended Uses: This OxHDL ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human OxHDL. This ELISA kit for research use only!

Principle of the Assay: OxHDL ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-OxHDL antibody and an OxHDL-HRP conjugate. The assay sample and buffer are incubated together with OxHDL-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the OxHDL concentration since OxHDL from samples and OxHDL-HRP conjugate compete for the anti-OxHDL antibody binding site. Since the number of sites is limited, as more sites are occupied by OxHDL from the sample, fewer sites are left to bind OxHDL-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The OxHDL concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/oxldl-rat-elisa-kits/oxidized-lowdensity-lipoprotein/86027>