

Choline phosphoglyceride, PC/CPG ELISA Kit | PC/CPG elisa kit

Fish choline phosphoglyceride, PC/CPG ELISA Kit

CATALOG #
AAA87985


PRODUCT SYNONYMS

Choline phosphoglyceride; PC/CPG; Fish choline phosphoglyceride; PC/CPG ELISA Kit; PC/CPG elisa kit

PRODUCT SPECIFICATIONS

Samples	Cell culture fluid & body fluid & tissue homogenate Serum or blood plasma
Preparation and Storage	Store all reagents at 2-8 degree C

IMAGES



ADDITIONAL INFORMATION

Related Product Information for PC/CPG elisa kit

For Samples: Cell culture fluid & body fluid & tissue homogenate Serum or blood plasma

Intended Uses: This PC/CPG ELISA kit is intended for laboratory research use only. The stop solution changes the color from blue to yellow and the intensity of the color is measured at 450 nm using a spectrophotometer. In order to measure the concentration of PC/CPG in the sample, this PC/CPG ELISA Kit includes a set of calibration standards. The calibration standards are assayed at the same time as the samples and allow the operator to produce a standard curve of Optical Density versus PC/CPG concentration. The concentration of in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Principle of the Assay: The coated well immunoenzymatic assay for the quantitative measurement of PC/CPG utilizes a multiclonal anti-PC/CPG antibody and an PC/CPG-HRP conjugate. The assay sample and buffer are incubated together with PC/CPG-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the PC/CPG concentration since PC/CPG from samples and PC/CPG-HRP conjugate compete for the anti-PC/CPG antibody binding site. Since the number of sites is limited, as more sites are occupied by PC/CPG from the sample, fewer sites are left to bind PC/CPG-HRP conjugate. Standards of known PC/CPG concentrations are run concurrently with the samples being assayed and a standard curve is plotted relating the intensity of the color (Optical Density) to the concentration of PC/CPG. The PC/CPG concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

