

Rat Undercarboxylated Osteocalcin ELISA Kit | ucOC elisa kit

Rat Undercarboxylated Osteocalcin ELISA Kit

CATALOG #
AAA88016

REACTIVITY
Rat


PRODUCT SYNONYMS

Undercarboxylated Osteocalcin; Rat Undercarboxylated Osteocalcin ELISA Kit; ucOC elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rat
Specificity	This assay has high sensitivity and excellent specificity for detection of ucOC. No significant cross-reactivity or interference between ucOC and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between ucOC and all the analogues, therefore, cross reaction may still exist in some cases.
Assay Type	Quantitative Competitive or Sandwich
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Sensitivity	0.1 ng/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for ucOC elisa kit	Principle of the Assay: ucOC ELISA kit applies the quantitative sandwich enzyme immunoassay technique. The microtiter plate has been pre-coated with a monoclonal antibody specific for ucOC. Standards or samples are then added to the microtiter plate wells and ucOC if present, will bind to the antibody pre-coated wells. In order to quantitatively determine the amount of ucOC present in the sample, a standardized preparation of horseradish peroxidase (HRP)-conjugated polyclonal antibody, specific for ucOC are added to each well to "sandwich" the ucOC immobilized on the plate. The microtiter plate undergoes incubation, and then the wells are thoroughly washed to remove all unbound components. Next, substrate solutions are added to each well. The enzyme (HRP) and substrate are allowed to react over a short incubation period. Only those wells that contain ucOC and enzyme-conjugated antibody will exhibit a change in color. The enzyme-substrate reaction is terminated by addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The ucOC concentration in each sample is interpolated from this standard curve. Intended Uses: This ucOC ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Rat ucOC. This ELISA kit for research use only, not for therapeutic or test applications!
---	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/ucoc-rat-elisa-kits/undercarboxylated-osteocalcin/88016>