

Goat Luteinizing Hormone ELISA Kit | LH elisa kit

Goat Luteinizing Hormone ELISA Kit

CATALOG #
AAA88351

REACTIVITY
Goat


PRODUCT SYNONYMS

Luteinizing Hormone; Goat Luteinizing Hormone ELISA Kit; LH elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Goat
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Sensitivity	0.1 ng/mL.
Preparation and Storage	Store all reagents at 2-8 degree C

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for LH elisa kit

For samples: Serum, plasma, cell culture supernatants, body fluid and tissue homogenate

INTENDED USE This ASM ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Rabbit ASM. This ELISA kit for research use only!

PRINCIPLE OF THE ASSAY ASM ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-ASM antibody and an ASM-HRP conjugate. The assay sample and buffer are incubated together with ASM-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the ASM concentration since ASM from samples and ASM-HRP conjugate compete for the anti-ASM antibody binding site. Since the number of sites is limited, as more sites are occupied by ASM from the sample, fewer sites are left to bind ASM-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The ASM concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/lh-goat-elisa-kits/luteinizing-hormone/88351>