

Porcine Vitamin B6 ELISA Kit | VB6 elisa kit

Porcine Vitamin B6 ELISA Kit

CATALOG #
AAA89194

REACTIVITY
Porcine


PRODUCT SYNONYMS

Vitamin B6; Porcine Vitamin B6 ELISA Kit; VB6 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Sensitivity	1.0 ng/mL.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for VB6 elisa kit

For samples: Serum, plasma, cell culture supernatants, body fluid and tissue homogenate

INTENDED USE This VB6 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Porcine VB6. This ELISA kit for research use only!

PRINCIPLE OF THE ASSAY VB6 ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-VB6 antibody and an VB6-HRP conjugate. The assay sample and buffer are incubated together with VB6-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the VB6 concentration since VB6 from samples and VB6-HRP conjugate compete for the anti-VB6 antibody binding site. Since the number of sites is limited, as more sites are occupied by VB6 from the sample, fewer sites are left to bind VB6-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The VB6 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/vb6-porcine-elisa-kits/vitamin-b6/89194>