

Canine Soluble endothelium-selectin ELISA Kit | SE-selectin elisa kit

Canine Soluble endothelium-selectin ELISA Kit

CATALOG #
AAA89392

REACTIVITY
Canine


PRODUCT SYNONYMS

Soluble endothelium-selectin; Canine Soluble endothelium-selectin ELISA Kit; SE-selectin elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	25-500pg/mL
Sensitivity	1.0pg/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for SE-selectin elisa kit

Principle of the Assay: SE-SELECTIN ELISA kit applies the quantitative sandwich enzyme immunoassay technique. The microtiter plate has been pre-coated with a monoclonal antibody specific for SE-SELECTIN. Standards or samples are then added to the microtiter plate wells and SE-SELECTIN if present, will bind to the antibody pre-coated wells. In order to quantitatively determine the amount of SE-SELECTIN present in the sample, a standardized preparation of horseradish peroxidase (HRP)-conjugated polyclonal antibody, specific for SE-SELECTIN are added to each well to "sandwich" the SE-SELECTIN immobilized on the plate. The microtiter plate undergoes incubation, and then the wells are thoroughly washed to remove all unbound components. Next, substrate solutions are added to each well. The enzyme (HRP) and substrate are allowed to react over a short incubation period. Only those wells that contain SE-SELECTIN and enzyme-conjugated antibody will exhibit a change in color. The enzyme-substrate reaction is terminated by addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The SE-SELECTIN concentration in each sample is interpolated from this standard curve.

Intended Uses: This SE-SELECTIN ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine SE-SELECTIN. This ELISA kit for research use only!

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/se-selectin-canine-elisa-kits/soluble-endothelium-selectin/89392>