

Porcine Carbohydrate antigen 19-9 ELISA Kit | CA19-9 elisa kit

Porcine Carbohydrate antigen 19-9 ELISA Kit

CATALOG #
AAA89777

REACTIVITY
Porcine


PRODUCT SYNONYMS

Carbohydrate antigen 19-9; Porcine Carbohydrate antigen 19-9 ELISA Kit; CA19-9 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Specificity	This assay has high sensitivity and excellent specificity for detection of CA19-9. No significant cross-reactivity or interference between CA19-9 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between CA19-9 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 U/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for CA19-9 elisa kit	Intended Uses: This CA19-9 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Porcine CA19-9. This ELISA kit for research use only! Principle of the Assay: CA19-9 ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-CA19-9 antibody and an CA19-9-HRP conjugate. The assay sample and buffer are incubated together with CA19-9-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the CA19-9 concentration since CA19-9 from samples and CA19-9-HRP conjugate compete for the anti-CA19-9 antibody binding site. Since the number of sites is limited, as more sites are occupied by CA19-9 from the sample, fewer sites are left to bind CA19-9-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The CA19-9 concentration in each sample is interpolated from this standard curve.
Product Categories/Family for CA19-9 elisa kit	Cancer

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/ca19-9-porcine-elisa-kits/carbohydrate-antigen-19-9/89777>