

Sheep Tight Junction Protein 1 ELISA Kit | TJP1 elisa kit

Sheep Tight Junction Protein 1 ELISA Kit

CATALOG #
AAA91054

REACTIVITY
Sheep


PRODUCT SYNONYMS

Tight Junction Protein 1; Sheep Tight Junction Protein 1 ELISA Kit; TJP1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Sheep
Specificity	This assay has high sensitivity and excellent specificity for detection of TJP-1. No significant cross-reactivity or interference between TJP-1 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between TJP-1 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	1.0 pg/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TJP1 elisa kit	Intended Uses: This TJP-1 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Sheep TJP-1. This ELISA kit for research use only! Principle of the Assay: TJP-1 ELISA kit applies the competitive enzyme immunoassay technique utilizing a polyclonal anti-TJP-1 antibody and an TJP-1-HRP conjugate. The assay sample and buffer are incubated together with TJP-1-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the TJP-1 concentration since TJP-1 from samples and TJP-1-HRP conjugate compete for the anti-TJP-1 antibody binding site. Since the number of sites is limited, as more sites are occupied by TJP-1 from the sample, fewer sites are left to bind TJP-1-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The TJP-1 concentration in each sample is interpolated from this standard curve.
--	---

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/tjp1-sheep-elisa-kits/tight-junction-protein-1/91054>