

Canine Microalbuminuria ELISA Kit | MAU elisa kit

Canine Microalbuminuria ELISA Kit

CATALOG #
AAA92025

REACTIVITY
Canine


PRODUCT SYNONYMS

Microalbuminuria; Canine Microalbuminuria ELISA Kit; MAU elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Sandwich
Detection Range	5.0-100mg/L
Sensitivity	1.0 mg/L
Intended Uses	This MAU ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine MAU. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MAU elisa kit

Principle of the assay: MAU ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-MAU antibody and an MAU-HRP conjugate. The assay sample and buffer are incubated together with MAU-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the MAU concentration since MAU from samples and MAU-HRP conjugate compete for the anti-MAU antibody binding site. Since the number of sites is limited, as more sites are occupied by MAU from the sample, fewer sites are left to bind MAU-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The MAU concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/mau-canine-elisa-kits/microalbuminuria/92025>