

Sheep Triiodothyronine ELISA Kit | T3 elisa kit

Sheep Triiodothyronine ELISA Kit

CATALOG #
AAA92557

REACTIVITY
Sheep


PRODUCT SYNONYMS

Triiodothyronine; Sheep Triiodothyronine ELISA Kit; T3 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Sheep
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Competitive (Quantitative)
Sensitivity	1.0 ng/mL.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for T3 elisa kit

Intended Uses: This T3 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Sheep T3. This ELISA kit for research use only!

Principle of the Assay: T3 ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-T3 antibody and an T3-HRP conjugate. The assay sample and buffer are incubated together with T3-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the T3 concentration since T3 from samples and T3-HRP conjugate compete for the anti-T3 antibody binding site. Since the number of sites is limited, as more sites are occupied by T3 from the sample, fewer sites are left to bind T3-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The T3 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.