

Guinea Pig Malondialdehyde ELISA Kit | MDA elisa kit

Guinea pig Malondialdehyde ELISA Kit

CATALOG #
AAA92640

REACTIVITY
Guinea Pig


PRODUCT SYNONYMS

Malondialdehyde; Guinea pig Malondialdehyde ELISA Kit; MDA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Guinea Pig
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	2.5-50ng/mL
Sensitivity	0.1ng/mL
Intended Uses	This MDA ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Guinea pig MDA. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MDA elisa kit

Principle of the assay: MDA ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-MDA antibody and an MDA-HRP conjugate. The assay sample and buffer are incubated together with MDA-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the MDA concentration since MDA from samples and MDA-HRP conjugate compete for the anti-MDA antibody binding site. Since the number of sites is limited, as more sites are occupied by MDA from the sample, fewer sites are left to bind MDA-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The MDA concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://stage.aaabiotech.com/mda-guinea-pig-pig-elisa-kits/malondialdehyde/92640>