

Bovine Anti-Methylglyoxal Antibody ELISA Kit | AMA elisa kit

Bovine Anti-Methylglyoxal Antibody ELISA Kit

CATALOG #
AAA93348

REACTIVITY
Bovine


PRODUCT SYNONYMS

Anti-Methylglyoxal Antibody; Bovine Anti-Methylglyoxal Antibody ELISA Kit; AMA elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Bovine
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Sandwich
Sensitivity	1.0 ng/mL.
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for AMA elisa kit

Intended Uses: This AMA ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Bovine AMA. This ELISA kit for research use only!

Principle of the Assay: AMA ELISA kit applies the quantitative sandwich enzyme immunoassay technique. The microtiter plate has been pre-coated with a monoclonal antibody specific for AMA. Standards or samples are then added to the microtiter plate wells and AMA if present, will bind to the antibody pre-coated wells. In order to quantitatively determine the amount of AMA present in the sample, a standardized preparation of horseradish peroxidase (HRP)-conjugated polyclonal antibody, specific for AMA are added to each well to "sandwich" the AMA immobilized on the plate. The microtiter plate undergoes incubation, and then the wells are thoroughly washed to remove all unbound components. Next, substrate solutions are added to each well. The enzyme (HRP) and substrate are allowed to react over a short incubation period. Only those wells that contain AMA and enzyme-conjugated antibody will exhibit a change in color. The enzyme-substrate reaction is terminated by addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The AMA concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.