

Porcine Bromodeoxyuridine ELISA Kit | BrdU elisa kit

Porcine Bromodeoxyuridine ELISA Kit

CATALOG #
AAA95279

REACTIVITY
Porcine


PRODUCT SYNONYMS

Bromodeoxyuridine; Porcine Bromodeoxyuridine ELISA Kit; BrdU elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Sensitivity	1.0 pmol/ml.
Intended Uses	This BrdU ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Porcine BrdU. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for BrdU elisa kit

Principle of the assay: BrdU ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-BrdU antibody and an BrdU-HRP conjugate. The assay sample and buffer are incubated together with BrdU-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the BrdU concentration since BrdU from samples and BrdU-HRP conjugate compete for the anti-BrdU antibody binding site. Since the number of sites is limited, as more sites are occupied by BrdU from the sample, fewer sites are left to bind BrdU-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The BrdU concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.