

## Canine Creatinine ELISA Kit | CREA elisa kit

Canine Creatinine ELISA Kit

CATALOG #  
**AAA95520**

REACTIVITY  
**Canine**


### PRODUCT SYNONYMS

Creatinine; Canine Creatinine ELISA Kit; CREA elisa kit

### PRODUCT SPECIFICATIONS

|                         |                                                                            |
|-------------------------|----------------------------------------------------------------------------|
| Reactivity              | Canine                                                                     |
| Samples                 | Serum, plasma, cell culture supernatants, body fluid and tissue homogenate |
| Assay Type              | Competitive                                                                |
| Sensitivity             | 1.0umol/L.                                                                 |
| Preparation and Storage | Store all reagents at 2-8 degree C.                                        |

### IMAGES



Standard Curve (Sample)

### ADDITIONAL INFORMATION

#### Related Product Information for CREA elisa kit

**Intended Uses:** This Creatinine ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine Creatinine. This ELISA kit for research use only!

**Principle of the Assay**||Creatinine ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-Creatinine antibody and an Creatinine-HRP conjugate. The assay sample and buffer are incubated together with Creatinine-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the Creatinine concentration since Creatinine from samples and Creatinine-HRP conjugate compete for the anti-Creatinine antibody binding site. Since the number of sites is limited, as more sites are occupied by Creatinine from the sample, fewer sites are left to bind Creatinine-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The Creatinine concentration in each sample is interpolated from this standard curve.

**For Research Use Only.** Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

