

Human Specificity protein 1 ELISA Kit | SP1 elisa kit

Human Specificity protein 1 ELISA Kit

CATALOG #
AAA97289

REACTIVITY
Human


PRODUCT SYNONYMS

Specificity protein 1; Human Specificity protein 1 ELISA Kit; SP1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive
Detection Range	25-500pg/mL
Sensitivity	1.0pg/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for SP1 elisa kit

Intended Uses: This SP1 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human SP1.

Principle of the Assay: SP1 ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-SP1 antibody and an SP1-HRP conjugate. The assay sample and buffer are incubated together with SP1-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the SP1 concentration since SP1 from samples and SP1-HRP conjugate compete for the anti-SP1 antibody binding site. Since the number of sites is limited, as more sites are occupied by SP1 from the sample, fewer sites are left to bind SP1-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The SP1 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

