

Human Glucose Transporter 2 ELISA Kit | GLUT2 elisa kit

Human Glucose Transporter 2 ELISA Kit

CATALOG #
AAA97666

REACTIVITY
Human


PRODUCT SYNONYMS

Glucose Transporter 2; Human Glucose Transporter 2 ELISA Kit; GLUT2/SLC2A2/GLUT-2/GLUT2 Glucose transporter type 2; liver/SLC2A2/solute carrier family 2 (facilitated glucose transporter); member 2/solute carrier family 2; facilitated glucose transporter member 2; GLUT2 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Sequence Length	395
Assay Type	Sandwich ELISA, Double Antibody
Samples	Serum, Plasma, Cell Culture Supernatant, cell or tissue lysate, Other liquid samples
Detection Range	0.625-40ng/ml
Sensitivity	0.375ng/ml
Preparation and Storage	Store entire kit at 2-8C for short-term. For longer-term, please store the microplate & standard at -20C, while the remaining reagents can be stored at 2-8C

IMAGES



Standard Curve (Sample)

Typical Data & Standard Curve

Results of a typical standard run of a Human GLUT2 ELISA Kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment. (N/A=not applicable)

X	ng/ml	0	0.625	1.25	2.5	5	10	20	40
Y	OD450	0.028	0.069	0.097	0.159	0.296	0.745	1.380	2.194

Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for GLUT2 elisa kit

Principle of the Assay||This kit was based on sandwich enzyme-linked immune-sorbent assay technology. anti-Human GLUT2 antibody was pre-coated onto 96-well plates. And the biotin conjugated anti-Human GLUT2 antibody was used as detection antibodies. The standards, test samples and biotin conjugated detection antibody were added to the wells subsequently, and wash with wash buffer. HRP-Streptavidin was added and unbound conjugates were washed away with wash buffer. TMB substrates were used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the Human GLUT2 amount of sample captured in plate. Read the O.D. absorbance at 450nm in a microplate reader, and then the concentration of Human GLUT2 can be calculated.

NCBI AND UNIPROT INFORMATION

NCBI GI #	270312991
Molecular Weight	42,867 Da
NCBI Official Full Name	glucose transporter 2, partial
UniProt Protein Name	Glucose transporter 2
UniProt Gene Name	glut2
UniProt Entry Name	D8L629_ORENI

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://stage.aaabiotech.com/glut2-human-elisa-kits/glucose-transporter-2/97666>